


Schedule of Accreditation

issued by

United Kingdom Accreditation Service

21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

 <p style="text-align: center;">Accredited to ISO/IEC 17025:2005</p>	<h3 style="margin: 0;">Lambda Calibration Ltd</h3> <p style="margin: 0;">Issue No: 029 Issue date: 21 April 2010</p>	
	<p>Lambda Calibration Ltd Units 11-13 Chorley Central Business Park Stump Lane Chorley Lancashire PR6 0BL</p>	<p>Contact: Brenda Seeds Tel: +44 (0)845 241 1533 Fax: +44 (0)845 241 1544 E-Mail: mail@lambda-cal.co.uk Website: www.lambda-cal.co.uk</p>
<p>Calibration performed by the Organisations at the locations specified below</p>		

Locations covered by the organisation and their relevant activities

Laboratory locations:

Location details	Activity	Location code		
<table style="width: 100%; border: none;"> <tr> <td style="width: 50%; border: none;">Address Lambda Calibration Ltd Units 11-13 Chorley Central Business Park Stump Lane Chorley Lancashire PR6 0BL</td> <td style="width: 50%; border: none;">Local contact Brenda Seeds</td> </tr> </table>	Address Lambda Calibration Ltd Units 11-13 Chorley Central Business Park Stump Lane Chorley Lancashire PR6 0BL	Local contact Brenda Seeds	<p>Dimensional Electrical Pressure Torque Temperature</p>	A
Address Lambda Calibration Ltd Units 11-13 Chorley Central Business Park Stump Lane Chorley Lancashire PR6 0BL	Local contact Brenda Seeds			

Site activities performed away from the locations listed above:

Location details	Activity	Location code		
<table style="width: 100%; border: none;"> <tr> <td style="width: 50%; border: none;">At customers premises</td> <td style="width: 50%; border: none;">Brenda Seeds</td> </tr> </table>	At customers premises	Brenda Seeds	<p>Dimensional Electrical</p>	B
At customers premises	Brenda Seeds			



0495
Accredited to
ISO/IEC 17025:2005

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

Lambda Calibration Ltd
Issue No: 029 Issue date: 21 April 2010

Calibration performed by the Organisation at the locations specified

DETAIL OF ACCREDITATION

Measured Quantity Instrument or Gauge	Range	Best Measurement Capability Expressed as an Expanded Uncertainty ($k=2$)	Remarks	Location Code
RANGE IN MILLIMETRES AND UNCERTAINTY IN MICROMETRES UNLESS OTHERWISE STATED				
LENGTH			NOTES	
Gauge blocks		Class (See Notes)	Class C uncertainties apply to the measurement of length of gauges by comparison with grade K standards of length of a similar material. Class C uncertainties apply to new and used grade 0, 1 and 2 gauges to BS 4311-1:2007 and BS EN ISO 3650:1999. 1. All calibrations must be carried out in accordance with procedures agreed by UKAS.	A
Inch (Steel and Carbide)	As BS 4311-1:2007 Up to 0.4 in Above 0.4 in up to 1 in Size 2 in Size 3 in Size 4 in	C 3 4 5 6 7		
Millimetre (Steel and Carbide)	As BS EN ISO 3650:1999 Up to 10 Above 10 up to 25 Sizes 30, 40, 50 60, 70, 75 80, 90, 100	C .08 .10 .12 .15 .18		
Gauge block accessories	As BS 4311-2	0.3		
Length gauges, flat and spherical ended (excluding Length Bars)	Up to 1 m	1 + (5 x length m)		
Plain plug gauges parallel, cylindrical setting standards and rollers.	From 1 up to 50 diameter above 50 up to 100 Above 100 up to 300	0.8 1 1.5		
Plain ring gauges (parallel)	From 2 up to 10 diameter Above 10 up to 50 Above 50 up to 100 Above 100 up to 300	1 0.8 1.0 2.5		
Precision balls (steel, carbide, ceramic)	From 1 up to 70 diameter	0.8		
Feeler gauges	As BS 957	1		
Gap gauges (Plain parallel)	As BS 969 From 2 up to 100 Above 100 up to 200 Above 200 up to 300	3 5 8		
Paint thickness setting foils	Up to 8	1.0		A
Rule - steel	As BS 4372 up to 1000	5 + (10 x length in m)		A



0495
Accredited to
ISO/IEC 17025:2005

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

Lambda Calibration Ltd
Issue No: 029 Issue date: 21 April 2010

Calibration performed by the Organisation at the locations specified

Measured Quantity Instrument or Gauge	Range	Best Measurement Capability Expressed as an Expanded Uncertainty ($k=2$)	Remarks	Location Code
RANGE IN MILLIMETRES AND UNCERTAINTY IN MICROMETRES UNLESS OTHERWISE STATED				
SCREW THREAD GAUGES			NOTES (cont'd)	
Screw plug gauges (parallel) including check and setting plugs. See Note 4	From 1 to 100 diameter Above 100 up to 150 diameter	3 on pitch 5 diameter	<p>2. In addition to all items other similar items, including parts of measuring instruments and machines, may be calibrated (see Note 1) to the uncertainties stated. Where the item or part calibrated is of lower quality due to wear, errors in geometry or from, poor surface texture, or where any other factor adversely affects the measurement capability, greater uncertainties must be quoted.</p> <p>3. The uncertainty quoted is for the departure from flatness, straightness, parallelism or squareness, i.e. the distance separating the two parallel planes which just enclose the surface under consideration</p> <p>4. Single start, symmetrical thread forms only.</p> <p>5. The uncertainty quoted is for the application of the calibration torque and does not take into account the characteristics of the device being calibrated.</p> <p>6. Calibrations may also be given in lbf.in and lbf.ft</p>	A
Screw plug gauges (taper) See Note 4	From 5 to 100 diameter Above 100 up to 150 diameter	5 on pitch 8 diameter		A
Screw ring gauges (parallel) See Note 4	From 1 mn up to 75 diameter Above 75 up to 150 diameter	5 on pitch 7 diameter		A
Screw ring gauges (taper) See Note 4	From 5 to 100 diameter Above 100 up to 150 diameter	7 on pitch 10 diameter Pitch: 1.5 Flank angle: $2+(800/(M \times P))$ Where M is projector magnification and P is pitch in mm		A
Screw thread adjustable caliper gauges (parallel) See Note 4	From 1 up to 150 diameter	By setting plugs		A
Thread measuring cylinders	As BS 5590 and specials	0.5		A
Parallels	As BS 906 up to 50 x 100 x 400	Dependent on size and grade From 1.5 up to 5		A
Vee blocks	As BS 3731 up to 200	Dependent on size and grade From 2.5 up to 5		A
Receiver and position gauges, jigs, fixtures	600 x 600 x 600	Dependent on size and features measured		A
Thread vee groove jaw blades	Down to 0.6 (40 T.P.I.)	3		A
ANGLE				
Squares				A
Blade type	As BS 939 up to 300 Above 300 up to 600	3 5	On squareness see Note 3	A
Cylindrical	As BS 939 up to 300 Above 300 up to 600	3 4		A
Block	As BS 939 up to 300 Above 300 up to 600	3 5		A



0495
Accredited to
ISO/IEC 17025:2005

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

Lambda Calibration Ltd
Issue No: 029 Issue date: 21 April 2010

Calibration performed by the Organisation at the locations specified

Measured Quantity Instrument or Gauge	Range	Best Measurement Capability Expressed as an Expanded Uncertainty ($k=2$)	Remarks	Location Code
RANGE IN MILLIMETRES AND UNCERTAINTY IN MICROMETRES UNLESS OTHERWISE STATED				
ANGLE (cont'd)				
Right angle and box angle plates	As BS 5535	Squareness: 3 + 1 per 100 mm Parallelism: 1 + 1 per 100 mm See Note 3		A
Sine bars and tables	As BS 3064 and up to 500 length	Linear dimensions 1 + (10 x length in m) Overall performance 3 seconds of arc		A
Clinometers	Up to 360°	10 seconds of arc or greater Dependant on sensitivity		A
Electronic indicating levels	8 minutes of arc	1% of range Minimum 0.4 seconds of arc		A
Levels, spirit	As BS 958:1968	1/5 div + 02 seconds of arc		A
MEASURING INSTRUMENTS				
Micrometers				
External Internal Depth	As BS 870 As BS 959 As BS 6468	Heads 2.0 Setting and extension rods 1 + (5 x length in m)		A
Micrometer, 3 point bore	From 6 up to 250	3 + (12 x length in m)		A
Micrometer Heads	As BS 1734	1.0		A
Height setting micrometer	Up to 300	Heads 1.5 between any two points Stepped column 2.5 Overall performance 3.0		A
Riser blocks for above	150 300	2.5 5		A
Vernier				A
Caliper gauges Height gauges Depth gauges	As BS 887 As BS 1643 up to 1.2 m As BS 6365	Overall performance 10 + (30 x length in m)		A
Bevel protractors	As BS 1685	1 min of arc + 1 vernier division		A



0495
Accredited to
ISO/IEC 17025:2005

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

Lambda Calibration Ltd
Issue No: 029 Issue date: 21 April 2010

Calibration performed by the Organisation at the locations specified

Measured Quantity Instrument or Gauge	Range	Best Measurement Capability Expressed as an Expanded Uncertainty ($k=2$)	Remarks	Location Code
MEASURING INSTRUMENTS				
Dial gauges and dial test indicators	As BS 907 and BS 2795	1.0		A
Displacement transducers (linear)	Up to 100	Dependant on range From 0.6 to 1.6		A
Height gauges (digital and electronic)	Up to 1000	2 + (10 x length in m)		A
FORM				
Surface plates				
Granite Cast Iron	BS 817	1.5 + (1.0 x diagonal in m)		A,B
ELECTRICAL MEASUREMENT			Published uncertainties apply for an environment of 18°C to 28°C. Measurements can take place outside these limits but at increased uncertainties.	
RESISTANCE	Up to 20 Ω 20 Ω to 200 Ω 200 Ω to 2 k Ω 2 k Ω to 20 k Ω 20 k Ω to 200 k Ω 200 k Ω to 2 M Ω 2 M Ω to 20 M Ω 20 M Ω to 200 M Ω 200 M Ω to 2 G Ω	23 ppm + 21 $\mu\Omega$ 15 ppm + 68 $\mu\Omega$ 12 ppm + 700 $\mu\Omega$ 12 ppm + 7 m Ω 14 ppm + 70 m Ω 26 ppm + 1.5 Ω 46 ppm + 4 Ω 360 ppm + 10 k Ω 0.36% + 900 k Ω		A & B
AC RESISTANCE	0.5 Ω to 2 k Ω 50 Hz	0.65%		A & B
DC VOLTS	Up to 200 mV 200 mV to 2 V 2 V to 20 V 20 V to 200 V 200 V to 1050 V	10.4 ppm + 62 nV 7.5 ppm + 482 nV 7.4 ppm + 2 μ V 12 ppm + 86 μ V 12 ppm + 256 μ V		A & B
DC CURRENT	1 kV to 10 kV	0.5%		
	Up to 200 μ A 200 μ A to 2 mA 2 mA to 20 mA 20 mA to 200 mA 200 mA to 1 A 1 A to 10 A 10 A to 1 kA	120 ppm + 0.5 nA 117 ppm + 5 nA 117 ppm + 48 nA 118 ppm + 1.2 μ A 234 ppm + 11 μ A 168 ppm + 13 μ A 356 ppm + 1.3 mA		A & B



0495
Accredited to
ISO/IEC 17025:2005

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

Lambda Calibration Ltd
Issue No: 029 Issue date: 21 April 2010

Calibration performed by the Organisation at the locations specified

Measured Quantity Instrument or Gauge	Range	Best Measurement Capability Expressed as an Expanded Uncertainty ($k=2$)	Remarks	Location Code
ELECTRICAL MEASUREMENT (cont'd)				
DC CURRENT Generation only	3.2 A to 105 A 105 A to 200 A 16 A to 160 A 160 A to 525 A 525 A to 1 KA	0.07% + 11 mA 0.07% + 50 mA 0.07% + 7 mA 0.07% +55 mA 0.07% + 250 mA	Suitable for Clamp Meters Using 10 turn coil Using 10 turn coil Using 50 turn coil Using 50 turn coil Using 50 turn coil	A & B
AC VOLTAGE	Up to 200 mV 20 Hz to 40 Hz 40 Hz to 10 kHz 10 kHz to 30 kHz 30 kHz to 100 kHz	300 ppm + 14 μ V 281 ppm + 2 μ V 363 ppm + 5 μ V 720 ppm + 20 μ V		A & B
	200 mV to 2 V 20 Hz to 40 Hz 40 Hz to 10 kHz 10 kHz to 30 kHz 30 kHz to 100 kHz	210 ppm + 120 μ V 164 ppm + 11 μ V 240 ppm + 30 μ V 496 ppm + 100 μ V		A & B
	2 V to 20 V 20 Hz to 40 Hz 40 Hz to 10 kHz 10 kHz to 30 kHz 30 kHz to 100 kHz	210 ppm + 1.2 mV 164 ppm + 0.1 mV 239 ppm + 0.3 mV 496 ppm + 1.0 mV		A & B
	20 V to 200 V 20 Hz to 40 Hz 40 Hz to 10 kHz 10 kHz to 30 kHz 30 kHz to 100 kHz	212 ppm + 13 mV 168 ppm + 1.6 mV 240 ppm + 4 mV 525 ppm + 11 mV		
	200 V to 1050 V 20 Hz to 40 Hz 40 Hz to 10 kHz 10 kHz to 30 kHz	212 ppm + 71 mV 168 ppm + 51 mV 340 ppm + 152 mV		
	1 kV to 10 kV @ 50 Hz	2.2%		
AC CURRENT	Up to 200 μ A 10 Hz to 1 kHz 1 kHz to 5 kHz	378 ppm + 21 nA 693 ppm + 21 nA		A & B
	200 μ A to 2 mA 10 Hz to 1 kHz 1 kHz to 5 kHz	378 ppm + 210 nA 693 ppm + 210 nA		
	2 mA to 20 mA 10 Hz to 1 kHz 1 kHz to 5 kHz	361 ppm + 2.1 μ A 693 ppm + 2.1 μ A		
	20 mA to 200 mA 10 Hz to 1 kHz 1 kHz to 5 kHz	378 ppm + 21 μ A 693 ppm + 21 μ A		



0495
Accredited to
ISO/IEC 17025:2005

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

Lambda Calibration Ltd
Issue No: 029 Issue date: 21 April 2010

Calibration performed by the Organisation at the locations specified

Measured Quantity Instrument or Gauge	Range	Best Measurement Capability Expressed as an Expanded Uncertainty ($k=2$)	Remarks	Location Code
ELECTRICAL MEASUREMENT (cont'd)				
AC CURRENT (cont'd)	200 mA to 1 A 10 Hz to 1 kHz 40 Hz to 3 kHz 3 kHz to 5 kHz	755 ppm + 209 μ A 0.12% 0.15%		
AC CURRENT Generation only	1 A to 10 A 40 Hz to 3 kHz 3 kHz to 5 kHz	0.16% 0.27%		
AC CURRENT Generation only	3.2 A to 32 A (10 to 100 Hz) 32 A to 200 A (10 to 100 Hz) 3.2 A to 32 A (100 to 440 Hz) 32 A to 200 A (100 to 440 Hz) 16 A to 160 A (10 to 100 Hz) 160 A to 1 kA (10 to 100 Hz)	0.33% + 6 mA 0.33% + 104 mA 0.93% + 31 mA 0.81% + 288 mA 0.33% + 323 mA 0.33% + 519 mA	Suitable for Clamp Meters Using 10 turn coil Using 10 turn coil Using 10 turn coil Using 50 turn coil	A & B
ELECTRICAL SIMULATION OF THERMOCOUPLES				
Type: B C E J K K L N R S T T	+500 °C to +1820 °C + 0 °C to +2320 °C -250 °C to +1000 °C -210 °C to +1200 °C -200 °C to -250 °C -200 °C to +1372 °C -200 °C to +900 °C -200 °C to +1300 °C + 0 °C to +1767 °C + 0 °C to +1767 °C -250 °C to -200 °C -200 °C to +400 °C	0.56 °C 0.42 °C 0.46 °C 0.27 °C 0.58 °C 0.29 °C 0.28 °C 0.34 °C 0.53 °C 0.50 °C 0.60 °C 0.29 °C		A & B
CAPACITANCE Generation only	Up to 350 Hz 0.5 nF to 4 nF 4 nF to 40 nF 40 nF to 400 nF 400 nF to 4 μ F 4 μ F to 40 μ F 40 μ F to 400 μ F 400 μ F to 4 mF 4 mF to 40 mF	0.35% + 18 pF 0.35% + 35 pF 0.35% + 185 pF 0.47% + 2 nF 0.57% + 19 nF 0.58% + 185 nF 0.58% + 2 μ F 1.16% + 69 μ F	Suitable for the testing of capacitance measuring devices	A & B
	350 Hz to 1.5 Hz 0.5 nF to 4 nF 4 nF to 40 nF 40 nF to 400 nF 400 nF to 4 μ F 4 μ F to 40 μ F 40 μ F to 400 μ F 400 μ F to 4 mF 4 mF to 40 mF	0.70% +35 pF 0.70% + 69 pF 0.70% + 370 pF 0.93% + 3 nF 1.16% + 37 nF 1.16% + 370 nF 1.16% + 4 μ F 2.31% +139 μ F		



0495
Accredited to
ISO/IEC 17025:2005

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

Lambda Calibration Ltd
Issue No: 029 Issue date: 21 April 2010

Calibration performed by the Organisation at the locations specified

Measured Quantity Instrument or Gauge	Range	Best Measurement Capability Expressed as an Expanded Uncertainty ($k=2$)	Remarks	Location Code
ELECTRICAL MEASUREMENT (cont'd)				
BANDWIDTH	3 dB point with respect to set point 10 Hz to 50 kHz 50 kHz to 100 MHz 100 MHz to 250 MHz	0.6% 2.4 % 5.2%	Appropriate for calibration of oscilloscopes	A & B
FREQUENCY	2 Hz to 250 MHz	2 in 10^9	May be expressed as time (1/f) for repetitive measurements	A
FREQUENCY Generation only	0.2 Hz to 250 MHz	0.29 ppm	May be expressed as time (1/f) for repetitive measurements	A & B
TIME Electronically Triggered				
Interval/Period Average	25 ns to 10 s	2 in 10^9 + (100 ns/No of periods)	Actual uncertainties quoted on certificate will include an allowance for the characteristics of the measured signal	A & B
Time Interval (A-A Event)	100 ns to 10^9 s	2 in 10^9 + 100 ns		A & B
Time Interval (A-B Event)	100 ns to 10^9 s	2 in 10^9 + 100 ns		
Time Interval	0.1 ms to 10 s	1.8 %	Appropriate for the calibration of RCD testers	A & B
TIME Mechanically Triggered	over 1 second	50 ms		A & B
TORQUE				
Hand torque tools (including drivers)	To BS EN ISO 6789:2003 0.2 N.m to 1000 N.m	1.5 % See Notes 5 & 6		A
PRESSURE				
<u>Hydraulic pressure (gauge)</u>				
Calibration of pressure indicating instruments and gauges	500 kPa to 7 MPa 7 MPa to 110 MPa 110 MPa to 400 MPa	0.011% 0.0092% 0.024%	Calibration of devices with an electrical output may be undertaken	A
Calibration of Piezoelectric pressure transducers at quasi- static pressures	500 kPa to 7 MPa 10 MPa to 400 MPa	0.35% 0.35%	Calibration of devices with a charge output may be undertaken	A



0495
Accredited to
ISO/IEC 17025:2005

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

Lambda Calibration Ltd
Issue No: 029 Issue date: 21 April 2010

Calibration performed by the Organisation at the locations specified

Measured Quantity Instrument or Gauge	Range	Best Measurement Capability Expressed as an Expanded Uncertainty ($k=2$)	Remarks	Location Code
<p>PRESSURE (cont'd)</p> <p><u>Gas pressure (gauge)</u></p> <p>Calibration of pressure indicating instruments and gauges</p>	<p>-95 kPa to -7 kPa -7 kPa to -3 kPa -3 kPa to -1.5 kPa -1.5 kPa to -150 Pa -150 Pa to -24 Pa -24 Pa to 24 Pa 24 Pa to 150 Pa 150 Pa to 1.5 kPa 1.5 kPa to 3 kPa 3 kPa to 7 kPa 7 kPa to 2 MPa 2 MPa to 6 MPa 6 MPa to 25 MPa</p>	<p>12 Pa 0.011 % + 0.9 Pa 0.011 % + 0.6 Pa 0.025 % + 0.4 Pa 0.22 % + 0.095 Pa 0.22 % + 0.045 Pa 0.22 % + 0.095 Pa 0.025 % + 0.4 Pa 0.011% + 0.6 Pa 0.011% + 0.9 Pa 0.0085 % 0.021 % + 1.4 kPa 0.021 % + 4.5 kPa</p>	<p>Absolute pressure calibrations can be undertaken using associated barometric pressure measurement correction. The uncertainties quoted will be increased by 12 Pa.</p>	A
<p><u>Gas pressure (absolute)</u></p> <p>Calibration of pressure indicating instruments and gauges</p>	<p>0.1 Pa to 1 Pa 1 Pa to 10 Pa 10 Pa to 100 Pa 100 Pa to 1 kPa 1 kPa to 3.5 kPa 3.5 kPa to 131 kPa 131 kPa to 200 kPa</p>	<p>1 % + 0.1 Pa 1 % + 0.3 Pa 1 % + 1.8 Pa 1 % + 18 Pa 80 Pa 12 Pa 80 Pa</p>		A
<p>Temperature</p> <p>Radiation thermometers (pyrometers)</p>	<p>20 to 100 °C 101 to 200 °C 201 to 300 °C 301 to 400 °C 401 to 500 °C</p>	<p>0.8degC 1.2degC 1.6degC 2.1degC 2.6degC</p>		A
END				